Quanterix
The Science of Precision Health

Tau 2.0 Calibrator Diluent
Certificate of Test

Product Number:	101631
Lot Number:	301011
Expiration:	10-Jan-2024
Storage:	2 - 8°C

Laboratory Analysis

Characteristic	Result	pH Target	PASS / FAIL
pH	7.33	7.30 - 7.50	PASS

Review/Approval — Sonia Khillan (Name) — QC Manager (Title) — Sonia Khillan 30AUG2023 (Signature/Date)

Quanterix Corporation
900 Middlesex Turnpike, Building 1, Billerica, MA 01821
techsupport@quanterix.com

©Quanterix Corp. Simoa® is a registered trademark of Quanterix Corp.
For research use only. Not for use in diagnositic procedures