Quanterix®

Discovery Fueled by Ultra-Sensitivity

Simoa® Sample Diluent A-E Kit

Certificate of Analysis

Product Number:	101477
Lot Number:	504415
Expiration:	20-May-2025
Storage:	2-8°C

Component	Part Number	Lot Number
Sample Diluent A	101472	407815
Sample Diluent B	101473	401622
Sample Diluent C	101474	425605
Sample Diluent D	101475	425610
Sample Diluent E [1]	101476	414115

Laboratory Analysis

Characteristic	Result	pH Target
pH Measurement	Pass	7.35 - 7.45
Characteristic	**Result**	**pH Target**
pH Measurement	Pass	8.95 - 9.05 [1]

[1] Sample Diluent E (101476) pH target differs from the all other components

Note: Result corresponds to each component

Review/Approval Dipesh Patel DOC Control II QA 30SEP2024

Name Title Signature/Date

Quanterix Corporation
900 Middlesex Turnpike, Building 1, Billerica, 01821
techsupport@quanterix.com

©2023 Quanterix Corp. Simoa® is a registered trademark of Quanterix Corp.
For research use only. Not for use in diagnostic procedures.