Quanterix®

Discovery Fueled by Ultra-Sensitivity

Simoa® p-Tau217 Advantage PLUS Reagent Kit

Certificate of Analysis

Product Number:	104588
Lot Number:	504443
Expiration:	22-May-2026
Platform(s):	HD-X

Component	Part Number	Lot Number
Bead Reagent	104572	432703
Detector Reagent	104573	433102
SBG Reagent	104574	433101
Sample Diluent Reagent	104575	432712
Calibrator Diluent Reagent	104577	432713
Control Diluent Reagent	104578	432714
Calibrator Concentrate	104584	329610
Control 1 Concentrate	104586	329612
Control 2 Concentrate	104587	329613
RGP 2.0 Reagent	104321	N/A[1]
Activation Buffer Reagent	104324	N/A[1]

[1] Reagents are not Kit Lot Specific

Data below represents results generated on the Simoa® HD-X Analyzer.

Release Materials*			
	Lot	Result (pg/mL)	Spec. Range (pg/mL)
Control 1 Concentrate	329612	0.094	0.067 - 0.124
Control 2 Concentrate	329613	2.38	1.75 - 3.25

*** Ranges shown are generated internally for new lot release only. Customer should generate their own control ranges.**

Calibrator Concentrate Reconstitution Volume (uL)	806
Control 1 Concentrate Reconstitution Volume (uL)	806
Control 2 Concentrate Reconstitution Volume (uL)	806

Calibration Curve



Calibrator Levels (pg/mL)	
A	0.000
B	0.00244
C	0.00977
D	0.0391
E	0.156
F	0.625
G	2.50
H	10.0

Review/Approval Dipesh Patel — Doc control II QA — 18 DEC 2024

Name — Title — Signature/Date

Quanterix Corporation
900 Middlesex Turnpike, Building 1, Billerica, 01821
techsupport@quanterix.com

©2024 Quanterix Corp. Simoa® is a registered trademark of Quanterix Corp.
For research use only. Not for use in diagnostic procedures.